

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042419\
 Data File : VX009150.D
 Acq On : 24 Apr 2019 18:49
 Operator : JC/SP
 Sample : K2555-06
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 HW-15

Quant Time: Apr 25 07:22:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	208633	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	344822	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	291361	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	119257	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.07	65	143377	48.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.52%	
35) Dibromofluoromethane	5.50	113	104262	47.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.22%	
50) Toluene-d8	8.71	98	421807	47.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.58%	
62) 4-Bromofluorobenzene	11.13	95	136761	43.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	87.80%	
Target Compounds						
16) Acetone	2.45	43	3896	2.555	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042419\
Data File : VX009150.D
Acq On : 24 Apr 2019 18:49
Operator : JC/SP
Sample : K2555-06
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
HW-15

Quant Time: Apr 25 07:22:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

