

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042420\
 Data File : VX015940.D
 Acq On : 24 Apr 2020 18:58
 Operator : JC/SP
 Sample : L2414-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS037MW042-200422

Quant Time: Apr 25 07:42:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 24 13:17:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	134085	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	257057	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	248265	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	122086	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	98474	52.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.02%	
35) Dibromofluoromethane	5.47	113	75122	53.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.16%	
50) Toluene-d8	8.70	98	323225	54.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.92%	
62) 4-Bromofluorobenzene	11.12	95	134556	55.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.78%	
Target Compounds						
3) Chloromethane	1.31	50	2251	1.482	ug/l	96
27) cis-1,2-Dichloroethene	4.57	96	52746	27.413	ug/l	89
44) Trichloroethene	7.19	130	17248	7.804	ug/l	99
64) Tetrachloroethene	9.32	164	10174	4.118	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042420\
Data File : VX015940.D
Acq On : 24 Apr 2020 18:58
Operator : JC/SP
Sample : L2414-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS037MW042-200422

Quant Time: Apr 25 07:42:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 24 13:17:59 2020
Response via : Initial Calibration

