

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042420\
 Data File : VX015942.D
 Acq On : 24 Apr 2020 19:44
 Operator : JC/SP
 Sample : L2414-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW081-200422

Quant Time: Apr 25 07:44:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 24 13:17:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.63	168	130267	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	250243	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	246533	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	120696	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	96106	52.25	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.50%	
35) Dibromofluoromethane	5.47	113	72307	52.98	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.96%	
50) Toluene-d8	8.70	98	315452	54.59	ug/l	0.00
Spiked Amount 50.000			Recovery	=	109.18%	
62) 4-Bromofluorobenzene	11.12	95	131683	56.19	ug/l	0.00
Spiked Amount 50.000			Recovery	=	112.38%	
Target Compounds						
3) Chloromethane	1.31	50	1372	0.930	ug/l	97
4) Vinyl Chloride	1.39	62	1073	0.568	ug/l	98
21) trans-1,2-Dichloroethene	3.15	96	2269	1.377	ug/l	93
27) cis-1,2-Dichloroethene	4.57	96	35642	19.067	ug/l	88
44) Trichloroethene	7.19	130	29681	13.796	ug/l	93
64) Tetrachloroethene	9.32	164	9455	3.854	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042420\
Data File : VX015942.D
Acq On : 24 Apr 2020 19:44
Operator : JC/SP
Sample : L2414-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW081-200422

Quant Time: Apr 25 07:44:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 24 13:17:59 2020
Response via : Initial Calibration

