

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042420\
 Data File : VX015943.D
 Acq On : 24 Apr 2020 20:07
 Operator : JC/SP
 Sample : L2414-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW024-200422

Quant Time: Apr 25 07:47:22 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 24 13:17:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	134789	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	258156	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	247126	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	117055	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	99270	52.16	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.32%	
35) Dibromofluoromethane	5.47	113	74801	53.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.26%	
50) Toluene-d8	8.70	98	319663	53.63	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.26%	
62) 4-Bromofluorobenzene	11.12	95	127729	52.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.66%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	2655	1.373	ug/l	82
44) Trichloroethene	7.19	130	9181	4.136	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042420\
Data File : VX015943.D
Acq On : 24 Apr 2020 20:07
Operator : JC/SP
Sample : L2414-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW024-200422

Quant Time: Apr 25 07:47:22 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 24 13:17:59 2020
Response via : Initial Calibration

