

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042420\
 Data File : VX015944.D
 Acq On : 24 Apr 2020 20:29
 Operator : JC/SP
 Sample : L2414-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ST006MW184-200422

Quant Time: Apr 25 07:49:24 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 24 13:17:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	137957	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	260306	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	251734	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.06	152	125133	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.03	65	98529	50.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.16%	
35) Dibromofluoromethane	5.47	113	76420	53.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.66%	
50) Toluene-d8	8.70	98	325621	54.18	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.36%	
62) 4-Bromofluorobenzene	11.12	95	136159	55.85	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.70%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	5961	3.011	ug/l	82
44) Trichloroethene	7.19	130	21562	9.634	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042420\
Data File : VX015944.D
Acq On : 24 Apr 2020 20:29
Operator : JC/SP
Sample : L2414-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ST006MW184-200422

Quant Time: Apr 25 07:49:24 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 24 13:17:59 2020
Response via : Initial Calibration

