

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042420\
 Data File : VX015950.D
 Acq On : 24 Apr 2020 22:45
 Operator : JC/SP
 Sample : L2415-05
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 28 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SS040MW041R-200423

Quant Time: Apr 25 08:01:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
 Quant Title : SW846 8260
 QLast Update : Fri Apr 24 13:17:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.64	168	133992	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.84	114	248284	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.10	117	245623	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.07	152	121969	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.04	65	95356	50.40	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.80%	
35) Dibromofluoromethane	5.47	113	73490	54.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	108.54%	
50) Toluene-d8	8.70	98	316844	55.27	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.54%	
62) 4-Bromofluorobenzene	11.12	95	130098	55.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	111.90%	
Target Compounds						
27) cis-1,2-Dichloroethene	4.57	96	8528	4.435	ug/l	85
44) Trichloroethene	7.20	130	3593	1.683	ug/l	98
64) Tetrachloroethene	9.32	164	1180	0.483	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042420\
Data File : VX015950.D
Acq On : 24 Apr 2020 22:45
Operator : JC/SP
Sample : L2415-05
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 28 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SS040MW041R-200423

Quant Time: Apr 25 08:01:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
Quant Title : SW846 8260
QLast Update : Fri Apr 24 13:17:59 2020
Response via : Initial Calibration

