


Data Path : Z:\voasrv\HPCHEM1\MSVOA X\Data\VX042420\
 Data File : VX015923.D
 Acq On : 24 Apr 2020 10:37
 Operator : JC/SP
 Sample : BFB
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X042420W.M
 Title : SW846 8260
 Last Update : Fri Apr 24 13:17:59 2020



AutoFind: Scans 1645, 1646, 1647; Background Corrected with Scan 1639

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.7	6848	PASS
75	95	30	60	50.3	18373	PASS
95	95	100	100	100.0	36538	PASS
96	95	5	9	6.6	2422	PASS
173	174	0.00	2	0.4	87	PASS
174	95	50	100	67.2	24557	PASS
175	174	5	9	7.0	1714	PASS
176	174	95	101	96.9	23792	PASS
177	176	5	9	6.5	1553	PASS