

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042425\
 Data File : VX045922.D
 Acq On : 24 Apr 2025 13:14
 Operator : JC/MD
 Sample : PB167707ZHE#03
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB167707ZHE#03

Quant Time: Apr 25 02:26:44 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	62997	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127004	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	120708	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	50725	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64481	55.971	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	111.940%
35) Dibromofluoromethane	5.379	113	45811	50.839	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.680%
50) Toluene-d8	8.647	98	161831	51.454	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.900%
62) 4-Bromofluorobenzene	11.079	95	60655	52.945	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	105.880%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	17079	36.103	ug/l	96
18) Methyl Acetate	2.709	43	2132	1.957	ug/l	94
20) Methylene Chloride	2.782	84	7887	8.905	ug/l	96
43) Isopropyl Acetate	6.348	43	11732	5.049	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042425\
Data File : VX045922.D
Acq On : 24 Apr 2025 13:14
Operator : JC/MD
Sample : PB167707ZHE#03
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB167707ZHE#03

Quant Time: Apr 25 02:26:44 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

