

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042425\
 Data File : VX045925.D
 Acq On : 24 Apr 2025 14:23
 Operator : JC/MD
 Sample : Q1852-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 ETGI-354

Quant Time: Apr 25 02:27:14 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	64343	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	127209	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	121330	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51727	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	64869	55.130	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.260%	
35) Dibromofluoromethane	5.379	113	46691	51.732	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.460%	
50) Toluene-d8	8.647	98	161809	51.364	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.720%	
62) 4-Bromofluorobenzene	11.079	95	61794	53.852	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	107.700%	
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	20595	42.625	ug/l	98
18) Methyl Acetate	2.703	43	2187	1.966	ug/l	98
20) Methylene Chloride	2.782	84	5823	6.437	ug/l #	88
43) Isopropyl Acetate	6.348	43	11575	4.973	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042425\
Data File : VX045925.D
Acq On : 24 Apr 2025 14:23
Operator : JC/MD
Sample : Q1852-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
ETGI-354

Quant Time: Apr 25 02:27:14 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

