

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042425\
 Data File : VX045928.D
 Acq On : 24 Apr 2025 15:33
 Operator : JC/MD
 Sample : Q1852-08
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 72-12013

Quant Time: Apr 25 02:27:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.543	168	66443	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	129208	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	125196	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	53533	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	67126	55.245	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	110.480%
35) Dibromofluoromethane	5.379	113	47968	52.325	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	104.640%
50) Toluene-d8	8.647	98	168299	52.597	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	105.200%
62) 4-Bromofluorobenzene	11.079	95	62973	54.030	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	108.060%
Target Compounds						
					Qvalue	
16) Acetone	2.379	43	21149	42.388	ug/l	96
18) Methyl Acetate	2.709	43	2415	2.102	ug/l #	88
20) Methylene Chloride	2.788	84	5440	5.824	ug/l	90
43) Isopropyl Acetate	6.348	43	11989	5.071	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042425\
Data File : VX045928.D
Acq On : 24 Apr 2025 15:33
Operator : JC/MD
Sample : Q1852-08
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
72-12013

Quant Time: Apr 25 02:27:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

