

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042519\
 Data File : VX009163.D
 Acq On : 25 Apr 2019 15:32
 Operator : JC/SP
 Sample : K2535-01
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-0016

Quant Time: Apr 26 10:12:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	205554	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	330404	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	283922	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	112752	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	138667	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
35) Dibromofluoromethane	5.50	113	100849	47.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.12%	
50) Toluene-d8	8.71	98	410625	48.56	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.12%	
62) 4-Bromofluorobenzene	11.13	95	131609	44.09	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.18%	
Target Compounds						
39) Methylcyclohexane	7.46	83	9148	2.199	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042519\
Data File : VX009163.D
Acq On : 25 Apr 2019 15:32
Operator : JC/SP
Sample : K2535-01
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-0016

Quant Time: Apr 26 10:12:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

