

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042519\
 Data File : VX009169.D
 Acq On : 25 Apr 2019 17:53
 Operator : JC/SP
 Sample : K2566-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-0014

Quant Time: Apr 26 10:21:58 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	201457	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	325491	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	281933	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	118145	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	6.06	65	135310	47.66	ug/l	0.00
Spiked Amount			Recovery	=	95.32%	
35) Dibromofluoromethane	5.50	113	98683	47.24	ug/l	0.00
Spiked Amount			Recovery	=	94.48%	
50) Toluene-d8	8.71	98	409666	49.18	ug/l	0.00
Spiked Amount			Recovery	=	98.36%	
62) 4-Bromofluorobenzene	11.13	95	135090	45.94	ug/l	0.00
Spiked Amount			Recovery	=	91.88%	


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042519\
Data File : VX009169.D
Acq On : 25 Apr 2019 17:53
Operator : JC/SP
Sample : K2566-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S37-0014

Quant Time: Apr 26 10:21:58 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 5.66 min Scan# 739

Delta R.T. -0.01 min

Lab File: VX009169.D

Acq: 25 Apr 2019 17:53

Instrument :

MSVOA_X

ClientSampleId :


S37-0014

Tot Ion:168 Resp: 201457

Ion Ratio Lower Upper

168 100

99 56.5 43.0 64.6



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VX0

5.66

80000



60000

40000

20000

0

Time--> 5.50 5.60 5.70 5.80



#33

1,2-Dichloroethane-d4

Concen: 47.659 ug/l

RT: 6.06 min Scan# 805

Delta R.T. -0.00 min

Lab File: VX009169.D


Acq: 25 Apr 2019 17:53

Tgt Ion: 65 Resp: 135310

Ion Ratio Lower Upper

65 100

67 53.7 0.0 108.6



Abundance Ion 64.90 (64.60 to 65.60): VX0

Ion 66.90 (66.60 to 67.60): VX0

6.06

60000

50000

40000


30000


20000

10000

0

Time--> 5.90 6.00 6.10 6.20









#34
 1,4-Difluorobenzene
 Concen: 50.000 ug/l
 RT: 6.86 min Scan# 936
 Delta R.T. -0.00 min
 Lab File: VX009169.D
 Acq: 25 Apr 2019 17:53

Instrument :
 MSVOA_X
 ClientSampleId :
 S37-0014

Tot Ion:114 Resp: 325491
 Ion Ratio Lower Upper
 114 100
 63 23.5 0.0 46.0
 88 16.9 0.0 32.6




Abundance Ion 113.90 (113.60 to 114.60): V
 Ion 63.00 (62.70 to 63.70): VX0
 Ion 87.90 (87.60 to 88.60): VX0






#35
 Dibromofluoromethane
 Concen: 47.241 ug/l
 RT: 5.50 min Scan# 713
 Delta R.T. -0.00 min
 Lab File: VX009169.D
 Acq: 25 Apr 2019 17:53

Tgt Ion:113 Resp: 98683
 Ion Ratio Lower Upper
 113 100
 111 103.8 82.9 124.3
 192 21.5 17.4 26.0



Abundance Ion 112.80 (112.50 to 113.50): V
 Ion 110.80 (110.50 to 111.50): V
 Ion 191.70 (191.40 to 192.40): V









#50
Toluene-d8
Concen: 49.175 ug/l
RT: 8.71 min Scan# 1240
Delta R.T. -0.00 min
Lab File: VX009169.D
Acq: 25 Apr 2019 17:53




Instrument :
MSVOA_X
ClientSampleId :
S37-0014


Tot Ion: 98 Resp: 409666
Ion Ratio Lower Upper
98 100
100 64.3 51.4 77.2



#62
4-Bromofluorobenzene
Concen: 45.937 ug/l
RT: 11.13 min Scan# 1637
Delta R.T. -0.00 min
Lab File: VX009169.D
Acq: 25 Apr 2019 17:53

Tgt Ion: 95 Resp: 135090
Ion Ratio Lower Upper
95 100
174 79.3 0.0 159.6
176 77.0 0.0 154.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.11 min Scan# 1469
Delta R.T. -0.00 min
Lab File: VX009169.D
Acq: 25 Apr 2019 17:53

Instrument :
MSVOA_X
ClientSampled :
S37-0014

Tot Ion:117 Resp: 281933
Ion Ratio Lower Upper
117 100
82 61.1 48.2 72.4
119 31.8 25.0 37.6




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): VX0
Ion 118.90 (118.60 to 119.60): V



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.08 min Scan# 1792
Delta R.T. -0.00 min
Lab File: VX009169.D
Acq: 25 Apr 2019 17:53

Tgt Ion:152 Resp: 118145
Ion Ratio Lower Upper
152 100
115 58.5 41.4 124.2
150 156.2 0.0 349.8



Abundance Ion 151.90 (151.60 to 152.60): V
Ion 114.90 (114.60 to 115.60): V
Ion 149.90 (149.60 to 150.60): V

