

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042522\
 Data File : VX028334.D
 Acq On : 25 Apr 2022 18:21
 Operator : JC/MD
 Sample : N2468-11MEDL 40X
 Misc : 4.90g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 WC-15-1-8MEDL

Quant Time: Apr 26 08:13:55 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	416990	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	710053	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	685331	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	344798	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	249189	45.832	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	91.660%
35) Dibromofluoromethane	5.391	113	224849	45.155	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	90.300%
50) Toluene-d8	8.653	98	842049	44.028	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	88.060%
62) 4-Bromofluorobenzene	11.085	95	318507	43.555	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	87.100%
Target Compounds						
					Qvalue	
39) Methylcyclohexane	7.379	83	7561	0.923	ug/l	93
84) 1,2,4-Trimethylbenzene	11.756	105	24056	1.203	ug/l	93
85) sec-Butylbenzene	11.890	105	29816	1.203	ug/l #	73

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042522\
Data File : VX028334.D
Acq On : 25 Apr 2022 18:21
Operator : JC/MD
Sample : N2468-11MEDL 40X
Misc : 4.90g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
WC-15-1-8MEDL

Quant Time: Apr 26 08:13:55 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

