

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042522\
 Data File : VX028316.D
 Acq On : 25 Apr 2022 10:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 26 08:08:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	122	0.00
2 T	Dichlorodifluoromethane	0.323	0.460	-42.4#	196#	0.00
3 P	Chloromethane	0.375	0.391	-4.3	147	0.00
4 C	Vinyl Chloride	0.504	0.474	6.0#	129	0.00
5 T	Bromomethane	0.522	0.475	9.0	119	0.00
6 T	Chloroethane	0.461	0.330	28.4#	113	0.00
7 T	Trichlorofluoromethane	1.112	0.886	20.3	108	0.00
8 T	Diethyl Ether	0.382	0.328	14.1	120	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.405	0.483	-19.3	181#	0.00
10 T	Methyl Iodide	0.509	0.633	-24.4	157#	0.00
11 T	Tert butyl alcohol	0.135	0.119	11.9	125	0.00
12 CM	1,1-Dichloroethene	0.384	0.435	-13.3#	170#	0.00
13 T	Acrolein	0.080	0.102	-27.5#	144	0.00
14 T	Allyl chloride	0.645	0.613	5.0	126	0.00
15 T	Acrylonitrile	0.266	0.256	3.8	122	0.00
16 T	Acetone	0.215	0.210	2.3	140	0.00
17 T	Carbon Disulfide	1.035	0.998	3.6	132	0.00
18 T	Methyl Acetate	0.580	0.567	2.2	134	0.00
19 T	Methyl tert-butyl Ether	1.602	1.605	-0.2	124	0.00
20 T	Methylene Chloride	0.484	0.505	-4.3	146	0.00
21 T	trans-1,2-Dichloroethene	0.523	0.500	4.4	123	0.00
22 T	Diisopropyl ether	1.469	1.374	6.5	119	0.00
23 T	Vinyl Acetate	1.283	1.217	5.1	117	0.00
24 P	1,1-Dichloroethane	0.867	0.830	4.3	129	0.00
25 T	2-Butanone	0.386	0.346	10.4	114	0.00
26 T	2,2-Dichloropropane	0.847	0.769	9.2	117	0.00
27 T	cis-1,2-Dichloroethene	0.623	0.598	4.0	124	0.00
28 T	Bromochloromethane	0.372	0.300	19.4	100	0.00
29 T	Tetrahydrofuran	0.252	0.220	12.7	112	0.00
30 C	Chloroform	1.041	0.948	8.9#	116	0.00
31 T	Cyclohexane	0.794	0.701	11.7	112	0.00
32 T	1,1,1-Trichloroethane	0.943	0.864	8.4	115	0.00
33 S	1,2-Dichloroethane-d4	0.652	0.562	13.8	109	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	115	0.00
35 S	Dibromofluoromethane	0.351	0.339	3.4	116	0.00
36 T	1,1-Dichloropropene	0.457	0.437	4.4	117	0.00
37 T	Ethyl Acetate	0.464	0.440	5.2	115	0.00
38 T	Carbon Tetrachloride	0.516	0.513	0.6	117	0.00
39 T	Methylcyclohexane	0.577	0.546	5.4	114	0.00
40 TM	Benzene	1.339	1.298	3.1	115	0.00
41 T	Methacrylonitrile	0.239	0.237	0.8	120	0.00
42 TM	1,2-Dichloroethane	0.497	0.468	5.8	114	0.00
43 T	Isopropyl Acetate	0.755	0.722	4.4	117	0.00
44 TM	Trichloroethene	0.388	0.402	-3.6	125	0.00
45 C	1,2-Dichloropropane	0.328	0.314	4.3#	117	0.00
46 T	Dibromomethane	0.257	0.255	0.8	121	0.00
47 T	Bromodichloromethane	0.494	0.483	2.2	115	0.00
48 T	Methyl methacrylate	0.347	0.332	4.3	117	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042522\
 Data File : VX028316.D
 Acq On : 25 Apr 2022 10:48
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_X
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 26 08:08:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.009	0.009	0.0	110	0.00
50 S	Toluene-d8	1.347	1.223	9.2	108	0.00
51 T	4-Methyl-2-Pentanone	0.485	0.459	5.4	115	0.00
52 CM	Toluene	0.930	0.887	4.6#	113	0.00
53 T	t-1,3-Dichloropropene	0.525	0.531	-1.1	115	0.00
54 T	cis-1,3-Dichloropropene	0.564	0.562	0.4	117	0.00
55 T	1,1,2-Trichloroethane	0.378	0.375	0.8	119	0.00
56 T	Ethyl methacrylate	0.548	0.556	-1.5	118	0.00
57 T	1,3-Dichloropropane	0.610	0.592	3.0	118	0.00
58 T	2-Chloroethyl Vinyl ether	0.257	0.262	-1.9	114	0.00
59 T	2-Hexanone	0.375	0.359	4.3	114	0.00
60 T	Dibromochloromethane	0.412	0.425	-3.2	118	0.00
61 T	1,2-Dibromoethane	0.415	0.413	0.5	115	0.00
62 S	4-Bromofluorobenzene	0.515	0.470	8.7	106	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	109	0.00
64 T	Tetrachloroethene	0.391	0.428	-9.5	129	0.00
65 PM	Chlorobenzene	1.066	1.068	-0.2	114	0.00
66 T	1,1,1,2-Tetrachloroethane	0.397	0.418	-5.3	119	0.00
67 C	Ethyl Benzene	1.788	1.826	-2.1#	113	0.00
68 T	m/p-Xylenes	0.728	0.750	-3.0	112	0.00
69 T	o-Xylene	0.725	0.734	-1.2	112	0.00
70 T	Styrene	1.175	1.238	-5.4	113	0.00
71 P	Bromoform	0.313	0.352	-12.5	122	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	105	0.00
73 T	Isopropylbenzene	3.437	3.436	0.0	113	0.00
74 T	N-amyl acetate	1.188	1.248	-5.1	118	0.00
75 P	1,1,2,2-Tetrachloroethane	1.115	1.115	0.0	114	0.00
76 T	1,2,3-Trichloropropane	0.984	1.033	-5.0	116	0.00
77 T	Bromobenzene	0.895	0.913	-2.0	117	0.00
78 T	n-propylbenzene	3.811	3.914	-2.7	112	0.00
79 T	2-Chlorotoluene	2.369	2.340	1.2	111	0.00
80 T	1,3,5-Trimethylbenzene	2.917	2.952	-1.2	110	0.00
81 T	trans-1,4-Dichloro-2-butene	0.331	0.346	-4.5	113	0.00
82 T	4-Chlorotoluene	2.729	2.712	0.6	110	0.00
83 T	tert-Butylbenzene	2.979	2.997	-0.6	111	0.00
84 T	1,2,4-Trimethylbenzene	2.900	2.964	-2.2	110	0.00
85 T	sec-Butylbenzene	3.594	3.675	-2.3	111	0.00
86 T	p-Isopropyltoluene	3.100	3.231	-4.2	111	0.00
87 T	1,3-Dichlorobenzene	1.682	1.743	-3.6	114	0.00
88 T	1,4-Dichlorobenzene	1.766	1.792	-1.5	114	0.00
89 T	n-Butylbenzene	2.546	2.679	-5.2	112	0.00
90 T	Hexachloroethane	0.498	0.539	-8.2	111	0.00
91 T	1,2-Dichlorobenzene	1.671	1.739	-4.1	115	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.248	0.247	0.4	110	0.00
93 T	1,2,4-Trichlorobenzene	0.994	1.128	-13.5	127	0.00
94 T	Hexachlorobutadiene	0.443	0.481	-8.6	126	0.00
95 T	Naphthalene	3.310	3.710	-12.1	119	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042522\
Data File : VX028316.D
Acq On : 25 Apr 2022 10:48
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_X
LabSampleId :
VSTDCCC050

Quant Time: Apr 26 08:08:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.975	1.116	-14.5	126	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6