



Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042522\
Data File : VX028315.D
Acq On : 25 Apr 2022 09:43
Operator : JC/MD
Sample : BFB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Title : SW846 8260
Last Update : Tue Apr 19 13:38:04 2022



AutoFind: Scans 1638, 1639, 1640; Background Corrected with Scan 1631

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.9	14708	PASS
75	95	30	60	50.8	41723	PASS
95	95	100	100	100.0	82184	PASS
96	95	5	9	6.9	5705	PASS
173	174	0.00	2	1.0	651	PASS
174	95	50	100	82.7	67981	PASS
175	174	5	9	7.4	5048	PASS
176	174	95	101	97.2	66069	PASS
177	176	5	9	6.6	4329	PASS