

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042523\
 Data File : VX035283.D
 Acq On : 25 Apr 2023 15:03
 Operator : JC/MD
 Sample : VX0425WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0425WBL01

Quant Time: Apr 25 16:42:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	246227	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	402667	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	359306	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	160276	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	145642	49.028	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.060%
35) Dibromofluoromethane	5.385	113	123651	50.032	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.060%
50) Toluene-d8	8.647	98	471152	50.536	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.080%
62) 4-Bromofluorobenzene	11.079	95	170565	48.633	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042523\
Data File : VX035283.D
Acq On : 25 Apr 2023 15:03
Operator : JC/MD
Sample : VX0425WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0425WBL01

Quant Time: Apr 25 16:42:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

