

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042523\
 Data File : VX035284.D
 Acq On : 25 Apr 2023 15:26
 Operator : JC/MD
 Sample : 02406-01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 SB-12-6.0-6.5

Quant Time: Apr 25 16:42:36 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	248962	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	411307	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	365781	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165392	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	151102	50.307	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	100.620%
35) Dibromofluoromethane	5.385	113	115427	45.723	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	91.440%
50) Toluene-d8	8.647	98	483486	50.769	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.540%
62) 4-Bromofluorobenzene	11.085	95	172921	48.269	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	96.540%
Target Compounds						
						Qvalue
20) Methylene Chloride	2.782	84	5694	1.904	ug/l	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042523\
Data File : VX035284.D
Acq On : 25 Apr 2023 15:26
Operator : JC/MD
Sample : 02406-01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
SB-12-6.0-6.5

Quant Time: Apr 25 16:42:36 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

