

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042525\
 Data File : VX045937.D
 Acq On : 25 Apr 2025 12:37
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 25 23:00:38 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 02 03:11:43 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	63947	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	126556	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.049	117	116710	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	51430	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	63083	53.944	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.880%
35) Dibromofluoromethane	5.379	113	46522	51.811	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.620%
50) Toluene-d8	8.647	98	156365	49.892	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.780%
62) 4-Bromofluorobenzene	11.079	95	59009	51.690	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	103.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042525\
Data File : VX045937.D
Acq On : 25 Apr 2025 12:37
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 25 23:00:38 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X040225W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 02 03:11:43 2025
Response via : Initial Calibration

