

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042619\
 Data File : VX009179.D
 Acq On : 26 Apr 2019 11:03
 Operator : JC/SP
 Sample : VX0426MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0426MBL01

Quant Time: Apr 27 04:15:39 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	202124	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.85	114	322079	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	274896	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	113193	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	132163	46.40	ug/l	0.00
Spiked Amount	50.000		Recovery	=	92.80%	
35) Dibromofluoromethane	5.49	113	99008	47.90	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.80%	
50) Toluene-d8	8.71	98	402672	48.85	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.70%	
62) 4-Bromofluorobenzene	11.13	95	127778	43.91	ug/l	0.00
Spiked Amount	50.000		Recovery	=	87.82%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042619\
Data File : VX009179.D
Acq On : 26 Apr 2019 11:03
Operator : JC/SP
Sample : VX0426MBL01
Misc : 5.0g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampled :
VX0426MBL01

Quant Time: Apr 27 04:15:39 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

