

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042619\
 Data File : VX009188.D
 Acq On : 26 Apr 2019 14:36
 Operator : JC/SP
 Sample : K2565-03
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 S81-ER-2019-1

Quant Time: Apr 27 05:18:04 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.66	168	215637	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	343436	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	296886	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	126745	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	140897	46.36	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.72%	
35) Dibromofluoromethane	5.50	113	104040	47.20	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.40%	
50) Toluene-d8	8.71	98	426664	48.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.08%	
62) 4-Bromofluorobenzene	11.13	95	142074	45.79	ug/l	0.00
Spiked Amount 50.000			Recovery	=	91.58%	
Target Compounds						
16) Acetone	2.45	43	3403	2.159	ug/l	89
20) Methylene Chloride	2.85	84	31642	11.741	ug/l	94
29) Tetrahydrofuran	5.13	42	117312	69.579	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042619\
Data File : VX009188.D
Acq On : 26 Apr 2019 14:36
Operator : JC/SP
Sample : K2565-03
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
S81-ER-2019-1

Quant Time: Apr 27 05:18:04 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

