

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042619\
 Data File : VX009191.D
 Acq On : 26 Apr 2019 15:47
 Operator : JC/SP
 Sample : K2591-02
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-5

Quant Time: Apr 27 05:35:06 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	208002	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	339243	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.12	117	293110	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	123058	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	139447	47.57	ug/l	0.00
Spiked Amount	50.000		Recovery	=	95.14%	
35) Dibromofluoromethane	5.50	113	103345	47.47	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.94%	
50) Toluene-d8	8.71	98	421895	48.59	ug/l	0.00
Spiked Amount	50.000		Recovery	=	97.18%	
62) 4-Bromofluorobenzene	11.13	95	140512	45.84	ug/l	0.00
Spiked Amount	50.000		Recovery	=	91.68%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042619\
Data File : VX009191.D
Acq On : 26 Apr 2019 15:47
Operator : JC/SP
Sample : K2591-02
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-WATER-REF-5

Quant Time: Apr 27 05:35:06 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

