

Data Path : Z:\VOASRV\HPCHEM1\MSVOA X\DATA\VX042619\
 Data File : VX009193.D
 Acq On : 26 Apr 2019 16:33
 Operator : JC/SP
 Sample : K2591-04
 Misc : 5.0mL/MSVOA X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 STORAGE-BLANK-SAMPLE-MANA

Quant Time: Apr 27 05:41:51 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
 Quant Title : SW846 8260
 QLast Update : Wed Apr 17 06:47:58 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	5.67	168	201945	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	6.86	114	329262	50.00	ug/l	0.00
63) Chlorobenzene-d5	10.11	117	280292	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.08	152	114438	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	6.06	65	134825	47.37	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.74%	
35) Dibromofluoromethane	5.50	113	98760	46.74	ug/l	0.00
Spiked Amount	50.000		Recovery	=	93.48%	
50) Toluene-d8	8.71	98	408026	48.42	ug/l	0.00
Spiked Amount	50.000		Recovery	=	96.84%	
62) 4-Bromofluorobenzene	11.13	95	132057	44.39	ug/l	0.00
Spiked Amount	50.000		Recovery	=	88.78%	

Target Compounds	Qvalue
------------------	--------

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA X\DATA\VX042619\
Data File : VX009193.D
Acq On : 26 Apr 2019 16:33
Operator : JC/SP
Sample : K2591-04
Misc : 5.0mL/MSVOA X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
STORAGE-BLANK-SAMPLE-MANA

Quant Time: Apr 27 05:41:51 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_X\METHOD\82X041519W.M
Quant Title : SW846 8260
QLast Update : Wed Apr 17 06:47:58 2019
Response via : Initial Calibration

