

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
 Data File : VX028346.D
 Acq On : 26 Apr 2022 13:53
 Operator : JC/MD
 Sample : N2553-01ME
 Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 370ME

Quant Time: Apr 27 02:04:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	520921	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	814760	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	711033	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	345479	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	266932	39.300	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	78.600%
35) Dibromofluoromethane	5.391	113	246011	43.055	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	86.120%
50) Toluene-d8	8.647	98	914268	41.660	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	83.320%
62) 4-Bromofluorobenzene	11.085	95	317170	37.798	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	75.600%
Target Compounds						
25) 2-Butanone	4.599	43	29421	7.314	ug/l	Qvalue 100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
Data File : VX028346.D
Acq On : 26 Apr 2022 13:53
Operator : JC/MD
Sample : N2553-01ME
Misc : 5.07g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
370ME

Quant Time: Apr 27 02:04:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

