

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
 Data File : VX028350.D
 Acq On : 26 Apr 2022 15:26
 Operator : JC/MD
 Sample : N2511-02
 Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 1527

Quant Time: Apr 27 02:05:03 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	494616	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	808457	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	742603	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	359739	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	269173	41.737	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	83.480%
35) Dibromofluoromethane	5.391	113	238168	42.008	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	84.020%
50) Toluene-d8	8.653	98	925644	42.508	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	85.020%
62) 4-Bromofluorobenzene	11.085	95	332395	39.921	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	79.840%
Target Compounds						
					Qvalue	
16) Acetone	2.416	43	22207	9.933	ug/l #	83
18) Methyl Acetate	2.727	43	10836	1.890	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
Data File : VX028350.D
Acq On : 26 Apr 2022 15:26
Operator : JC/MD
Sample : N2511-02
Misc : 5.01g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
1527

Quant Time: Apr 27 02:05:03 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

