

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
 Data File : VX028354.D
 Acq On : 26 Apr 2022 16:59
 Operator : JC/MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 IBLK

Quant Time: Apr 27 02:06:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	529857	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	848564	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	801599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	404112	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	324760	47.007	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	94.020%
35) Dibromofluoromethane	5.391	113	307738	51.713	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	103.420%
50) Toluene-d8	8.653	98	1119430	48.977	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	97.960%
62) 4-Bromofluorobenzene	11.079	95	422450	48.339	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	96.680%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
Data File : VX028354.D
Acq On : 26 Apr 2022 16:59
Operator : JC/MD
Sample : IBLK
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
IBLK

Quant Time: Apr 27 02:06:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

