

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
 Data File : VX028355.D
 Acq On : 26 Apr 2022 17:22
 Operator : JC/MD
 Sample : PB144365TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144365TB

Quant Time: Apr 27 02:06:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	519218	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	851602	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	804284	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	390848	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	333420	49.250	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.500%
35) Dibromofluoromethane	5.391	113	314125	52.598	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.200%
50) Toluene-d8	8.653	98	1155590	50.379	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	100.760%
62) 4-Bromofluorobenzene	11.079	95	427169	48.705	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.400%
Target Compounds						
					Qvalue	
16) Acetone	2.380	43	13715	5.266	ug/l	94
20) Methylene Chloride	2.788	84	23508	4.680	ug/l	98
43) Isopropyl Acetate	6.342	43	32467	2.524	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
Data File : VX028355.D
Acq On : 26 Apr 2022 17:22
Operator : JC/MD
Sample : PB144365TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144365TB

Quant Time: Apr 27 02:06:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

