

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
 Data File : VX028356.D
 Acq On : 26 Apr 2022 17:46
 Operator : JC/MD
 Sample : PB144365ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144365ZHE#01

Quant Time: Apr 27 02:06:54 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	475454	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	796134	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	767154	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	378836	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	323041	52.109	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 104.220%			
35) Dibromofluoromethane	5.391	113	299713	53.681	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 107.360%			
50) Toluene-d8	8.653	98	1105579	51.556	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.120%			
62) 4-Bromofluorobenzene	11.079	95	411353	50.169	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 100.340%			
Target Compounds					Qvalue	
16) Acetone	2.380	43	12275	5.115	ug/l	95
20) Methylene Chloride	2.788	84	22328	4.854	ug/l	96
43) Isopropyl Acetate	6.342	43	32923	2.737	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
Data File : VX028356.D
Acq On : 26 Apr 2022 17:46
Operator : JC/MD
Sample : PB144365ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144365ZHE#01

Quant Time: Apr 27 02:06:54 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

