

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
 Data File : VX028359.D
 Acq On : 26 Apr 2022 18:55
 Operator : JC/MD
 Sample : PB144366TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144366TB

Quant Time: Apr 27 02:07:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	454837	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	763903	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	759687	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	365028	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	308089	51.949	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.900%
35) Dibromofluoromethane	5.391	113	282303	52.696	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	105.400%
50) Toluene-d8	8.653	98	1058931	51.465	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.920%
62) 4-Bromofluorobenzene	11.085	95	395875	50.319	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	100.640%
Target Compounds						Qvalue
16) Acetone	2.380	43	13100	5.869	ug/l	93
20) Methylene Chloride	2.788	84	21617	4.913	ug/l	98
43) Isopropyl Acetate	6.348	43	32809	2.843	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
Data File : VX028359.D
Acq On : 26 Apr 2022 18:55
Operator : JC/MD
Sample : PB144366TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144366TB

Quant Time: Apr 27 02:07:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

