

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
 Data File : VX028361.D
 Acq On : 26 Apr 2022 19:42
 Operator : JC/MD
 Sample : N2559-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-10S

Quant Time: Apr 27 02:08:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	450904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	770118	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	771482	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	379189	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	317882	54.068	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.140%	
35) Dibromofluoromethane	5.391	113	289942	53.685	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.380%	
50) Toluene-d8	8.653	98	1091107	52.600	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.200%	
62) 4-Bromofluorobenzene	11.085	95	405591	51.138	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.280%	
Target Compounds						Qvalue
16) Acetone	2.380	43	23297	11.643	ug/l	94
20) Methylene Chloride	2.788	84	21327	4.889	ug/l	97
43) Isopropyl Acetate	6.342	43	30075	2.585	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
Data File : VX028361.D
Acq On : 26 Apr 2022 19:42
Operator : JC/MD
Sample : N2559-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-10S

Quant Time: Apr 27 02:08:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

