

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
 Data File : VX028362.D
 Acq On : 26 Apr 2022 20:05
 Operator : JC/MD
 Sample : N2559-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-10D

Quant Time: Apr 27 02:08:25 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	434641	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	744298	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	741696	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	357682	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	310454	54.781	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	109.560%
35) Dibromofluoromethane	5.391	113	282437	54.110	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	108.220%
50) Toluene-d8	8.653	98	1054405	52.594	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.180%
62) 4-Bromofluorobenzene	11.079	95	393047	51.275	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.560%
Target Compounds						
						Qvalue
16) Acetone	2.386	43	21680	11.191	ug/l	# 85
20) Methylene Chloride	2.794	84	21376	5.084	ug/l	97
43) Isopropyl Acetate	6.342	43	30500	2.713	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042622\
Data File : VX028362.D
Acq On : 26 Apr 2022 20:05
Operator : JC/MD
Sample : N2559-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-10D

Quant Time: Apr 27 02:08:25 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

