

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042623\
 Data File : VX035316.D
 Acq On : 26 Apr 2023 16:20
 Operator : JC/MD
 Sample : 02419-09ME
 Misc : 5.01g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 EW924ME

Quant Time: Apr 27 04:43:01 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041823WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Apr 25 03:03:54 2023
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By : John Carlone 04/27/2023
 Supervised By : Mahesh Dadoda 05/01/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	6.757	114	472498	50.000	ug/L	0.00
28) Chlorobenzene-d5	10.055	117	406454	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	12.024	152	199851	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.367	65	70219m	23.048	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	46.100%#
7) Chloroethane-d5	1.672	69	49730m	22.589	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	45.180%#
11) 1,1-Dichloroethene-d2	2.288	65	39083	26.810	ug/L	-0.02
Spiked Amount	50.000	Range	60 - 125	Recovery	=	53.620%#
21) 2-Butanone-d5	4.483	46	142451	70.163	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	70.160%
24) Chloroform-d	5.056	84	179987	31.370	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	62.740%#
26) 1,2-Dichloroethane-d4	5.952	65	126838	34.425	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	68.860%#
32) Benzene-d6	5.964	84	399683	34.232	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	68.460%#
36) 1,2-Dichloropropane-d6	7.305	67	126670	35.516	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	71.040%
41) Toluene-d8	8.647	98	364943	34.200	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	68.400%#
43) trans-1,3-Dichloroprop...	8.951	79	43445	27.874	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	55.740%#
47) 2-Hexanone-d5	9.415	63	104842m	73.202	ug/L	0.04
Spiked Amount	100.000	Range	45 - 130	Recovery	=	73.200%
56) 1,1,2,2-Tetrachloroeth...	11.195	84	152739	34.969	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	69.940%
66) 1,2-Dichlorobenzene-d4	12.323	152	141168	36.450	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	72.900%#

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042623\
Data File : VX035316.D
Acq On : 26 Apr 2023 16:20
Operator : JC/MD
Sample : 02419-09ME
Misc : 5.01g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
EW924ME

Quant Time: Apr 27 04:43:01 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\SFAMXML041823WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Apr 25 03:03:54 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By : John Carlone 04/27/2023
Supervised By : Mahesh Dadoda 05/01/2023

