

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042624\
 Data File : VX041251.D
 Acq On : 26 Apr 2024 12:15
 Operator : JC/MD
 Sample : PB160528TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160528TB

Quant Time: Apr 27 02:15:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	202095	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	262778	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	227872	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.018	152	115229	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	98822	51.368	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	102.740%
35) Dibromofluoromethane	5.385	113	92845	58.084	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	116.160%
50) Toluene-d8	8.647	98	273829	50.775	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.560%
62) 4-Bromofluorobenzene	11.079	95	103141	51.815	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.640%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	39769	56.792	ug/l	94
18) Methyl Acetate	2.709	43	2280	1.101	ug/l	# 90
20) Methylene Chloride	2.782	84	6239	3.648	ug/l	# 91
37) Ethyl Acetate	4.739	43	2874	1.281	ug/l	# 91
43) Isopropyl Acetate	6.342	43	98256	27.182	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042624\
Data File : VX041251.D
Acq On : 26 Apr 2024 12:15
Operator : JC/MD
Sample : PB160528TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160528TB

Quant Time: Apr 27 02:15:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration

