

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042624\
 Data File : VX041252.D
 Acq On : 26 Apr 2024 12:37
 Operator : JC/MD
 Sample : PB160528ZHE#11
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160528ZHE#11

Quant Time: Apr 27 02:15:48 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	203650	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.757	114	237597	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	219912	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	108858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	110320	56.907	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 113.820%		
35) Dibromofluoromethane	5.385	113	93098	64.415	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 128.840%#		
50) Toluene-d8	8.647	98	270991	55.574	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 111.140%		
62) 4-Bromofluorobenzene	11.079	95	96636	53.693	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 107.380%		
Target Compounds						
						Qvalue
16) Acetone	2.386	43	42321	59.975	ug/l	90
18) Methyl Acetate	2.709	43	2274	1.089	ug/l	99
20) Methylene Chloride	2.782	84	4148	2.407	ug/l	94
37) Ethyl Acetate	4.739	43	2511	1.238	ug/l #	93
43) Isopropyl Acetate	6.342	43	124610	38.126	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042624\
Data File : VX041252.D
Acq On : 26 Apr 2024 12:37
Operator : JC/MD
Sample : PB160528ZHE#11
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160528ZHE#11

Quant Time: Apr 27 02:15:48 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration

