

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042624\
 Data File : VX041255.D
 Acq On : 26 Apr 2024 13:46
 Operator : JC/MD
 Sample : PB160528ZHE#14
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160528ZHE#14

Quant Time: Apr 27 02:17:11 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.549	168	163602	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	238488	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	234549	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	114128	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	91354	58.659	ug/l	0.00
Spiked Amount	50.000	Range 78 - 117	Recovery	= 117.320%#		
35) Dibromofluoromethane	5.385	113	77430	53.374	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 106.740%		
50) Toluene-d8	8.646	98	282674	57.754	ug/l	0.00
Spiked Amount	50.000	Range 92 - 112	Recovery	= 115.500%#		
62) 4-Bromofluorobenzene	11.079	95	103609	57.352	ug/l	0.00
Spiked Amount	50.000	Range 83 - 123	Recovery	= 114.700%		
Target Compounds						
					Qvalue	
16) Acetone	2.385	43	33270	58.690	ug/l	96
18) Methyl Acetate	2.715	43	2413	1.439	ug/l	94
20) Methylene Chloride	2.794	84	3226	2.330	ug/l	89
43) Isopropyl Acetate	6.342	43	96894	29.536	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042624\
Data File : VX041255.D
Acq On : 26 Apr 2024 13:46
Operator : JC/MD
Sample : PB160528ZHE#14
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160528ZHE#14

Quant Time: Apr 27 02:17:11 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration

