

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042624\
 Data File : VX041257.D
 Acq On : 26 Apr 2024 14:32
 Operator : JC/MD
 Sample : PB160522ZHE#02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB160522ZHE#02

Quant Time: Apr 27 02:18:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 23 05:30:28 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	5.550	168	136503	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	6.763	114	202278	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	198749	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	101705	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	75447	58.063	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	116.120%
35) Dibromofluoromethane	5.385	113	63287	51.435	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.860%
50) Toluene-d8	8.647	98	235750	56.789	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	113.580%#
62) 4-Bromofluorobenzene	11.079	95	89290	58.273	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	116.540%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	30845	65.214	ug/l	98
18) Methyl Acetate	2.703	43	2426	1.734	ug/l	95
20) Methylene Chloride	2.782	84	5090	4.406	ug/l #	85
43) Isopropyl Acetate	6.342	43	85691	30.797	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042624\
Data File : VX041257.D
Acq On : 26 Apr 2024 14:32
Operator : JC/MD
Sample : PB160522ZHE#02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB160522ZHE#02

Quant Time: Apr 27 02:18:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042224W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 23 05:30:28 2024
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 5.550 min Scan# 71
Delta R.T. -0.000 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32




Instrument :
MSVOA_X
ClientSampleId :
PB160522ZHE#02


Tgt Ion:168 Resp: 136503
Ion Ratio Lower Upper
168 100
99 47.0 56.6 85.0#



#16
Acetone
Concen: 65.214 ug/l
RT: 2.386 min Scan# 213
Delta R.T. -0.012 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32

Tgt Ion: 43 Resp: 30845
Ion Ratio Lower Upper
43 100
58 28.2 21.9 32.9









#18
Methyl Acetate
Concen: 1.734 ug/l
RT: 2.703 min Scan# 218
Delta R.T. -0.006 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32




Instrument :
MSVOA_X
ClientSampleId :
PB160522ZHE#02


Tgt Ion: 43 Resp: 2426
Ion Ratio Lower Upper
43 100
74 24.5 17.8 26.8



#20
Methylene Chloride
Concen: 4.406 ug/l
RT: 2.782 min Scan# 278
Delta R.T. -0.006 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32

Tgt Ion: 84 Resp: 5090
Ion Ratio Lower Upper
84 100
49 110.9 103.6 155.4
51 33.4 33.8 50.6#
86 50.1 49.4 74.0





#33

1,2-Dichloroethane-d4

Concen: 58.063 ug/l

RT: 5.952 min Scan# 798

Delta R.T. -0.000 min

Lab File: VX041257.D

Acq: 26 Apr 2024 14:32

Instrument :

MSVOA_X

ClientSampleId :





PB160522ZHE#02

Tgt Ion: 65 Resp: 75447

Ion Ratio Lower Upper

65 100

67 50.4 0.0 100.2



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 6.763 min Scan# 931

Delta R.T. 0.006 min

Lab File: VX041257.D

Acq: 26 Apr 2024 14:32




Tgt Ion: 114 Resp: 202278


Ion Ratio Lower Upper

114 100

63 18.7 0.0 45.0

88 14.7 0.0 36.6









#35
Dibromofluoromethane
Concen: 51.435 ug/l
RT: 5.385 min Scan# 705
Delta R.T. -0.000 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32




Instrument :
MSVOA_X
ClientSampleId :
PB160522ZHE#02


Tgt Ion:113 Resp: 63287
Ion Ratio Lower Upper
113 100
111 103.5 82.9 124.3
192 26.1 13.3 19.9#



#43
Isopropyl Acetate
Concen: 30.797 ug/l
RT: 6.342 min Scan# 862
Delta R.T. -0.006 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32

Tgt Ion: 43 Resp: 85691
Ion Ratio Lower Upper
43 100
61 19.1 15.2 22.8
87 13.7 9.8 14.6









#50
Toluene-d8
Concen: 56.789 ug/l
RT: 8.647 min Scan# 11
Delta R.T. -0.000 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32




Instrument :
MSVOA_X
ClientSampleId :
PB160522ZHE#02


Tgt Ion: 98 Resp: 235750
Ion Ratio Lower Upper
98 100
100 65.7 53.1 79.7



#62
4-Bromofluorobenzene
Concen: 58.273 ug/l
RT: 11.079 min Scan# 1639
Delta R.T. -0.000 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32

Tgt Ion: 95 Resp: 89290
Ion Ratio Lower Upper
95 100
174 100.6 0.0 135.6
176 98.0 0.0 131.0









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 10.055 min Scan# 1471
Delta R.T. -0.000 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32

Instrument :
MSVOA_X
ClientSampleId :
PB160522ZHE#02

Tgt Ion:117 Resp: 198749
Ion Ratio Lower Upper
117 100
82 50.2 46.2 69.4
119 32.7 25.5 38.3



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 12.024 min Scan# 1794
Delta R.T. -0.000 min
Lab File: VX041257.D
Acq: 26 Apr 2024 14:32

Tgt Ion:152 Resp: 101705
Ion Ratio Lower Upper
152 100
115 51.8 43.3 129.9
150 155.5 0.0 345.0

