

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
 Data File : VX028368.D
 Acq On : 27 Apr 2022 11:07
 Operator : JC/MD
 Sample : VX0427WBL01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 VX0427WBL01

Quant Time: Apr 27 15:52:17 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	418795	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	711275	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	687202	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	330855	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	295150	54.051	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 108.100%	
35) Dibromofluoromethane	5.391	113	269234	53.975	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 107.960%	
50) Toluene-d8	8.653	98	983849	51.354	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 102.700%	
62) 4-Bromofluorobenzene	11.079	95	360811	49.255	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 98.520%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
Data File : VX028368.D
Acq On : 27 Apr 2022 11:07
Operator : JC/MD
Sample : VX0427WBL01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
VX0427WBL01

Quant Time: Apr 27 15:52:17 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

