

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
 Data File : VX028372.D
 Acq On : 27 Apr 2022 12:50
 Operator : JC/MD
 Sample : N2508-01
 Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 40576-SOL

Quant Time: Apr 27 15:54:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	451792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	749541	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	682780	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	325415	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.964	65	254175	43.147	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	86.300%
35) Dibromofluoromethane	5.391	113	223093	42.442	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	84.880%
50) Toluene-d8	8.647	98	863266	42.759	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	85.520%
62) 4-Bromofluorobenzene	11.085	95	303642	39.335	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	78.660%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
Data File : VX028372.D
Acq On : 27 Apr 2022 12:50
Operator : JC/MD
Sample : N2508-01
Misc : 5.08g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
40576-SOL

Quant Time: Apr 27 15:54:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

