

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
 Data File : VX028375.D
 Acq On : 27 Apr 2022 14:00
 Operator : JC/MD
 Sample : PB144400TB
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144400TB

Quant Time: Apr 28 04:22:02 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	457753	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	761982	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	747138	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	359767	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	307462	51.514	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	103.020%
35) Dibromofluoromethane	5.391	113	279331	52.273	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	104.540%
50) Toluene-d8	8.653	98	1045885	50.959	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.920%
62) 4-Bromofluorobenzene	11.079	95	387539	49.383	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	98.760%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	15412	7.098	ug/l	99
20) Methylene Chloride	2.788	84	21852	4.934	ug/l	96
43) Isopropyl Acetate	6.342	43	30366	2.638	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
Data File : VX028375.D
Acq On : 27 Apr 2022 14:00
Operator : JC/MD
Sample : PB144400TB
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144400TB

Quant Time: Apr 28 04:22:02 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

