

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
 Data File : VX028376.D
 Acq On : 27 Apr 2022 14:23
 Operator : JC/MD
 Sample : PB144400ZHE#01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144400ZHE#01

Quant Time: Apr 28 04:22:22 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	434587	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	743810	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	743714	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	355696	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	310366	54.772	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 109.540%	
35) Dibromofluoromethane	5.391	113	275523	52.820	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 105.640%	
50) Toluene-d8	8.653	98	1045873	52.203	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 104.400%	
62) 4-Bromofluorobenzene	11.079	95	390304	50.951	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.900%	
Target Compounds						
						Qvalue
16) Acetone	2.380	43	14915	7.262	ug/l	98
20) Methylene Chloride	2.788	84	21427	5.096	ug/l	94
43) Isopropyl Acetate	6.342	43	29943	2.665	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
Data File : VX028376.D
Acq On : 27 Apr 2022 14:23
Operator : JC/MD
Sample : PB144400ZHE#01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144400ZHE#01

Quant Time: Apr 28 04:22:22 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

