

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
 Data File : VX028379.D
 Acq On : 27 Apr 2022 15:33
 Operator : JC/MD
 Sample : PB144390ZHE#04
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144390ZHE#04

Quant Time: Apr 28 04:23:14 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	435385	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	749985	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	752056	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	367086	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	312399	55.030	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.060%
35) Dibromofluoromethane	5.385	113	283719	53.943	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.880%
50) Toluene-d8	8.653	98	1068760	52.906	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.820%
62) 4-Bromofluorobenzene	11.079	95	399512	51.723	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.440%
Target Compounds						
						Qvalue
16) Acetone	2.379	43	14140	6.797	ug/l	96
20) Methylene Chloride	2.788	84	22102	5.247	ug/l	98
43) Isopropyl Acetate	6.342	43	29774	2.628	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
Data File : VX028379.D
Acq On : 27 Apr 2022 15:33
Operator : JC/MD
Sample : PB144390ZHE#04
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144390ZHE#04

Quant Time: Apr 28 04:23:14 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

