

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
 Data File : VX028380.D
 Acq On : 27 Apr 2022 15:56
 Operator : JC/MD
 Sample : PB144390ZHE#05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PB144390ZHE#05

Quant Time: Apr 28 04:23:33 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	425152	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	745070	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	748632	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	353708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	305435	55.098	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	110.200%
35) Dibromofluoromethane	5.391	113	279432	53.479	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	106.960%
50) Toluene-d8	8.653	98	1055690	52.604	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	105.200%
62) 4-Bromofluorobenzene	11.079	95	389476	50.757	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	101.520%
Target Compounds						
						Qvalue
16) Acetone	2.380	43	14312	7.096	ug/l	90
20) Methylene Chloride	2.788	84	23294	5.663	ug/l	96
43) Isopropyl Acetate	6.342	43	29562	2.626	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
Data File : VX028380.D
Acq On : 27 Apr 2022 15:56
Operator : JC/MD
Sample : PB144390ZHE#05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PB144390ZHE#05

Quant Time: Apr 28 04:23:33 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

