

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
 Data File : VX028382.D
 Acq On : 27 Apr 2022 16:42
 Operator : JC/MD
 Sample : N2572-02
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 MH-14S

Quant Time: Apr 28 04:24:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	437242	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	754490	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	758713	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	362262	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	317423	55.677	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	111.360%
35) Dibromofluoromethane	5.391	113	285375	53.934	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	107.860%
50) Toluene-d8	8.653	98	1080496	53.168	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	106.340%
62) 4-Bromofluorobenzene	11.079	95	401778	51.706	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.420%
Target Compounds						Qvalue
16) Acetone	2.379	43	18609	9.343	ug/l	99
20) Methylene Chloride	2.788	84	19872	4.698	ug/l	97
43) Isopropyl Acetate	6.348	43	29058	2.549	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
Data File : VX028382.D
Acq On : 27 Apr 2022 16:42
Operator : JC/MD
Sample : N2572-02
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
MH-14S

Quant Time: Apr 28 04:24:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

