

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
 Data File : VX028385.D
 Acq On : 27 Apr 2022 17:52
 Operator : JC/MD
 Sample : N2577-07ME
 Misc : 7.33g/5mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GT-04-(25-30)ME

Quant Time: Apr 28 04:25:35 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 19 13:38:04 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	492778	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	801667	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	722627	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	355669	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	268118	41.729	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery =	83.460%		
35) Dibromofluoromethane	5.391	113	242300	43.098	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery =	86.200%		
50) Toluene-d8	8.653	98	916027	42.422	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery =	84.840%		
62) 4-Bromofluorobenzene	11.086	95	323109	39.135	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery =	78.260%		
Target Compounds						
16) Acetone	2.368	43	245	Below Cal	Qvalue #	39

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042722\
Data File : VX028385.D
Acq On : 27 Apr 2022 17:52
Operator : JC/MD
Sample : N2577-07ME
Misc : 7.33g/5mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GT-04-(25-30)ME

Quant Time: Apr 28 04:25:35 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X041922W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 19 13:38:04 2022
Response via : Initial Calibration

