

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042722\  
 Data File : VX028383.D  
 Acq On : 27 Apr 2022 17:06  
 Operator : JC/MD  
 Sample : N2572-05  
 Misc : 5.0mL/MSVOA\_X/WATER  
 ALS Vial : 19 Sample Multiplier: 1

Instrument :  
 MSVOA\_X  
 ClientSampleId :  
 MH-14D

Quant Time: Apr 28 04:24:32 2022  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
 Quant Title : SW846 8260  
 QLast Update : Tue Apr 19 13:38:04 2022  
 Response via : Initial Calibration


| Compound                    | R.T. QIon |       | Response | Conc     | Units | Dev(Min) |
|-----------------------------|-----------|-------|----------|----------|-------|----------|
| -----                       |           |       |          |          |       |          |
| Internal Standards          |           |       |          |          |       |          |
| 1) Pentafluorobenzene       | 5.556     | 168   | 432704   | 50.000   | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene     | 6.763     | 114   | 756744   | 50.000   | ug/l  | 0.00     |
| 63) Chlorobenzene-d5        | 10.055    | 117   | 753861   | 50.000   | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4  | 12.024    | 152   | 358283   | 50.000   | ug/l  | 0.00     |
| System Monitoring Compounds |           |       |          |          |       |          |
| 33) 1,2-Dichloroethane-d4   | 5.958     | 65    | 309202   | 54.804   | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range | 61 - 141 | Recovery | =     | 109.600% |
| 35) Dibromofluoromethane    | 5.391     | 113   | 276572   | 52.115   | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range | 69 - 133 | Recovery | =     | 104.220% |
| 50) Toluene-d8              | 8.653     | 98    | 1044799  | 51.258   | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range | 65 - 126 | Recovery | =     | 102.520% |
| 62) 4-Bromofluorobenzene    | 11.079    | 95    | 386524   | 49.595   | ug/l  | 0.00     |
| Spiked Amount               | 50.000    | Range | 58 - 135 | Recovery | =     | 99.180%  |
| Target Compounds            |           |       |          |          |       |          |
|                             |           |       |          |          |       | Qvalue   |
| 16) Acetone                 | 2.379     | 43    | 28517    | 15.238   | ug/l  | 94       |
| 20) Methylene Chloride      | 2.788     | 84    | 19783    | 4.726    | ug/l  | 97       |
| 43) Isopropyl Acetate       | 6.348     | 43    | 28300    | 2.476    | ug/l  | 96       |
| -----                       |           |       |          |          |       |          |


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_X\Data\VX042722\  
Data File : VX028383.D  
Acq On : 27 Apr 2022 17:06  
Operator : JC/MD  
Sample : N2572-05  
Misc : 5.0mL/MSVOA\_X/WATER  
ALS Vial : 19 Sample Multiplier: 1

Instrument :  
MSVOA\_X  
ClientSampleId :  
MH-14D


Quant Time: Apr 28 04:24:32 2022  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_X\Method\82X041922W.M  
Quant Title : SW846 8260  
QLast Update : Tue Apr 19 13:38:04 2022  
Response via : Initial Calibration








#1  
Pentafluorobenzene  
Concen: 50.000 ug/l  
RT: 5.556 min Scan# 733  
Delta R.T. -0.000 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06

Instrument :  
MSVOA\_X  
ClientSampleId :  
MH-14D







Tgt Ion:168 Resp: 432704  
Ion Ratio Lower Upper  
168 100  
99 54.7 41.7 62.5



#16  
Acetone  
Concen: 15.238 ug/l  
RT: 2.379 min Scan# 212  
Delta R.T. -0.001 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06


Tgt Ion: 43 Resp: 28517  
Ion Ratio Lower Upper  
43 100  
58 32.3 28.8 43.2







#20  
Methylene Chloride  
Concen: 4.726 ug/l  
RT: 2.788 min Scan# 21  
Delta R.T. -0.000 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06

Instrument :  
MSVOA\_X  
ClientSampleId :  
MH-14D



Tgt Ion: 84 Resp: 19783




| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 84  | 100   |       |       |
| 49  | 111.4 | 85.5  | 128.3 |
| 51  | 35.0  | 26.6  | 39.8  |
| 86  | 65.2  | 51.5  | 77.3  |




#33  
1,2-Dichloroethane-d4  
Concen: 54.804 ug/l  
RT: 5.958 min Scan# 799  
Delta R.T. 0.000 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06

Tgt Ion: 65 Resp: 309202

| Ion | Ratio | Lower | Upper |
|-----|-------|-------|-------|
| 65  | 100   |       |       |
| 67  | 51.2  | 0.0   | 107.4 |









#34  
1,4-Difluorobenzene  
Concen: 50.000 ug/l  
RT: 6.763 min Scan# 91  
Delta R.T. -0.000 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06




Instrument :  
MSVOA\_X  
ClientSampleId :  
MH-14D

Tgt Ion:114 Resp: 756744  
Ion Ratio Lower Upper  
114 100  
63 18.2 0.0 36.8  
88 15.4 0.0 32.4



#35  
Dibromofluoromethane  
Concen: 52.115 ug/l  
RT: 5.391 min Scan# 706  
Delta R.T. 0.006 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06

Tgt Ion:113 Resp: 276572  
Ion Ratio Lower Upper  
113 100  
111 103.0 82.8 124.2  
192 20.3 16.9 25.3








#43  
Isopropyl Acetate  
Concen: 2.476 ug/l  
RT: 6.348 min Scan# 863  
Delta R.T. 0.012 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06




Instrument :  
MSVOA\_X  
ClientSampleId :  
MH-14D


Tgt Ion: 43 Resp: 28300  
Ion Ratio Lower Upper  
43 100  
61 20.2 18.5 27.7  
87 15.3 13.0 19.6



#50  
Toluene-d8  
Concen: 51.258 ug/l  
RT: 8.653 min Scan# 1241  
Delta R.T. -0.000 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06

Tgt Ion: 98 Resp: 1044799  
Ion Ratio Lower Upper  
98 100  
100 66.2 52.6 78.8









#62  
4-Bromofluorobenzene  
Concen: 49.595 ug/l  
RT: 11.079 min Scan# 1639  
Delta R.T. 0.000 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06




Instrument :  
MSVOA\_X  
ClientSampleId :  
MH-14D


Tgt Ion: 95 Resp: 386524  
Ion Ratio Lower Upper  
95 100  
174 87.8 0.0 163.0  
176 84.2 0.0 155.0



#63  
Chlorobenzene-d5  
Concen: 50.000 ug/l  
RT: 10.055 min Scan# 1471  
Delta R.T. -0.000 min  
Lab File: VX028383.D  
Acq: 27 Apr 2022 17:06

Tgt Ion: 117 Resp: 753861  
Ion Ratio Lower Upper  
117 100  
82 52.7 41.8 62.8  
119 32.6 24.8 37.2





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 12.024 min Scan# 11

Delta R.T. 0.000 min

Lab File: VX028383.D


Acq: 27 Apr 2022 17:06

Instrument :

MSVOA\_X

ClientSampleId :

MH-14D



Tgt Ion:152 Resp: 358283

Ion Ratio Lower Upper

152 100

115 57.5 37.5 112.7

150 156.9 0.0 338.2

