

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035333.D
 Acq On : 27 Apr 2023 12:06
 Operator : JC/MD
 Sample : 02445-02ME
 Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TRE-1730-1731

Quant Time: Apr 27 16:24:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.544	168	253626	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.757	114	416864	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	372681	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171351	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	150312	49.124	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.240%
35) Dibromofluoromethane	5.385	113	115873	45.288	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	90.580%
50) Toluene-d8	8.647	98	492942	51.072	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	102.140%
62) 4-Bromofluorobenzene	11.085	95	176963	48.739	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.480%
Target Compounds						
					Qvalue	
20) Methylene Chloride	2.782	84	4456	1.462	ug/l #	81
68) m/p-Xylenes	10.305	106	1639	0.306	ug/l	87
84) 1,2,4-Trimethylbenzene	11.756	105	2428	0.229	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035333.D
Acq On : 27 Apr 2023 12:06
Operator : JC/MD
Sample : 02445-02ME
Misc : 5.03g/10mL/100uL/5.00mL/MSVOA_X/MEOH
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TRE-1730-1731

Quant Time: Apr 27 16:24:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

