

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035337.D
 Acq On : 27 Apr 2023 13:38
 Operator : JC/MD
 Sample : 02496-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 TOTE-4-25-2023

Quant Time: Apr 28 05:04:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	242318	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	397322	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	357977	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	165544	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	142092	48.605	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.200%
35) Dibromofluoromethane	5.385	113	118741	48.691	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.380%
50) Toluene-d8	8.647	98	464226	50.463	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	100.920%
62) 4-Bromofluorobenzene	11.079	95	174463	50.414	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	100.820%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	37738	29.722	ug/l	93
68) m/p-Xylenes	10.305	106	1613	0.314	ug/l	85
95) Naphthalene	13.780	128	6003	0.543	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035337.D
Acq On : 27 Apr 2023 13:38
Operator : JC/MD
Sample : 02496-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
TOTE-4-25-2023

Quant Time: Apr 28 05:04:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

