

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035343.D
 Acq On : 27 Apr 2023 15:58
 Operator : JC/MD
 Sample : 02492-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 22-MW-04I

Quant Time: Apr 28 05:06:54 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	232086	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	379620	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	341723	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	152828	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	138474	49.455	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	98.920%
35) Dibromofluoromethane	5.391	113	117198	50.300	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.600%
50) Toluene-d8	8.647	98	447909	50.959	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.920%
62) 4-Bromofluorobenzene	11.079	95	163558	49.466	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.940%
Target Compounds						
					Qvalue	
4) Vinyl Chloride	1.373	62	49971	22.302	ug/l	100
12) 1,1-Dichloroethene	2.325	96	4498	1.995	ug/l	90
21) trans-1,2-Dichloroethene	3.093	96	13563	5.540	ug/l	98
27) cis-1,2-Dichloroethene	4.489	96	2886116	1002.921	ug/l	86
30) Chloroform	5.092	83	2248	0.504	ug/l #	71
44) Trichloroethene	7.141	130	10189675	3544.109	ug/l	98
64) Tetrachloroethene	9.274	164	26069	9.963	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035343.D
Acq On : 27 Apr 2023 15:58
Operator : JC/MD
Sample : 02492-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 18 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
22-MW-04I

Quant Time: Apr 28 05:06:54 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

