

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035344.D
 Acq On : 27 Apr 2023 16:21
 Operator : JC/MD
 Sample : 02495-01
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 PZ-01

Quant Time: Apr 28 05:07:16 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.550	168	214453	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	351754	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	321900	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	149693	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	131529	50.837	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 101.680%	
35) Dibromofluoromethane	5.385	113	107943	49.998	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 100.000%	
50) Toluene-d8	8.647	98	417105	51.214	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 102.420%	
62) 4-Bromofluorobenzene	11.079	95	155679	50.813	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 101.620%	
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	17749	15.795	ug/l	92
27) cis-1,2-Dichloroethene	4.495	96	4290	1.613	ug/l	87
44) Trichloroethene	7.129	130	24443	9.175	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035344.D
Acq On : 27 Apr 2023 16:21
Operator : JC/MD
Sample : 02495-01
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
PZ-01

Quant Time: Apr 28 05:07:16 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

