

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035346.D
 Acq On : 27 Apr 2023 17:07
 Operator : JC/MD
 Sample : 02495-05
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GMW-11

Quant Time: Apr 28 05:07:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	230002	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	372529	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	333179	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	150175	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.958	65	135873	48.966	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	97.940%
35) Dibromofluoromethane	5.385	113	114786	50.202	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.400%
50) Toluene-d8	8.647	98	437483	50.721	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.440%
62) 4-Bromofluorobenzene	11.079	95	158040	48.707	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	97.420%
Target Compounds						
					Qvalue	
16) Acetone	2.386	43	2343	1.944	ug/l	94
27) cis-1,2-Dichloroethene	4.507	96	768	0.269	ug/l	70
40) Benzene	6.044	78	2121	0.207	ug/l #	67
44) Trichloroethene	7.129	130	4674	1.657	ug/l	95
95) Naphthalene	13.774	128	217584	21.679	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035346.D
Acq On : 27 Apr 2023 17:07
Operator : JC/MD
Sample : 02495-05
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GMW-11

Quant Time: Apr 28 05:07:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

