

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
 Data File : VX035347.D
 Acq On : 27 Apr 2023 17:31
 Operator : JC/MD
 Sample : 02495-07
 Misc : 5.0mL/MSVOA_X/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_X
 ClientSampleId :
 GMW-X

Quant Time: Apr 28 05:08:12 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
 Quant Title : SW846 8260
 QLast Update : Tue Apr 25 16:31:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	5.556	168	240189	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	6.763	114	393329	50.000	ug/l	0.00
63) Chlorobenzene-d5	10.055	117	362096	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	12.024	152	171887	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	5.952	65	144043	49.709	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	=	99.420%
35) Dibromofluoromethane	5.385	113	121911	50.499	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	101.000%
50) Toluene-d8	8.647	98	460685	50.586	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	=	101.180%
62) 4-Bromofluorobenzene	11.079	95	177400	51.783	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.560%
Target Compounds						
					Qvalue	
16) Acetone	2.392	43	3041	2.416	ug/l	95
17) Carbon Disulfide	2.514	76	2353	0.421	ug/l #	88
40) Benzene	6.037	78	2207	0.204	ug/l #	82
44) Trichloroethene	7.135	130	3024	1.015	ug/l	96
95) Naphthalene	13.774	128	219336	19.093	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_X\Data\VX042723\
Data File : VX035347.D
Acq On : 27 Apr 2023 17:31
Operator : JC/MD
Sample : 02495-07
Misc : 5.0mL/MSVOA_X/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_X
ClientSampleId :
GMW-X

Quant Time: Apr 28 05:08:12 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_X\Method\82X042523W.M
Quant Title : SW846 8260
QLast Update : Tue Apr 25 16:31:40 2023
Response via : Initial Calibration

